

***Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program***

130526

**C**

**TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.**

## DIAGNOSTIC CHECK

[illegible]

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUL 06 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130526 ✓
A	Name	JOSEPH V. SHOEMAKER	ID# 4188 Date 7-5-09 ✓
	Agency	(SDPI) SAND POINT PD	907 383-3700
	Instrument Location	SAND POINT, AK	99661 ✓
B	Alco S/N	X173004 ✓	Target Value .079 ✓ High Pressure 1100
	Alco Test Value Average	.078 ✓	.078 ✓
		1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco
	Signature	Joseph V. Shoemaker ✓	
		CJB 7/17/09	
(OVER)			

I, Nita J. Bolz, after being first duly sworn, depose and state the following (Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 21<sup>st</sup> day of Jul 2009.

Carolyn M. Noland  
Notary Seal Stamp  
Notary Public, State of Alaska  
Commission Expires with Office

